

FRONT DETECTOR ()

SIGNAL 1

Data rate: 20 Hz
Type: test plot
Save Data: Off
Zero: 0.0 (Off)
Range: 0
Fast Peaks: Off
Attenuation: 0

COLUMN COMP 1

(No Detectors Installed)

THERMAL AUX 2

Use: MSD Transfer Line Heater
Description:
Initial temp: 300 'C (On)
Initial time: 0.00 min
Rate Final temp Final time
1 0.0(Off)

BACK DETECTOR ()

SIGNAL 2

Data rate: 20 Hz
Type: test plot
Save Data: Off
Zero: 0.0 (Off)
Range: 0
Fast Peaks: Off
Attenuation: 0

COLUMN COMP 2

(No Detectors Installed)

POST RUN

Post Time: 0.00 min

TIME TABLE

Time Specifier Parameter & Setpoint

7673 Injector

Front Injector:

Sample Washes 0
Sample Pumps 3
Injection Volume 1.0 microliters
Syringe Size 10.0 microliters
PostInj Solvent A Washes 3
PostInj Solvent B Washes 3
Viscosity Delay 3 seconds
Plunger Speed Fast
PreInjection Dwell 0.00 minutes
PostInjection Dwell 0.00 minutes

Back Injector:

0 parameters specified

Column 1 Inventory Number : ?

Column 2 Inventory Number :

MS ACQUISITION PARAMETERS

General Information

Tune File : atune.u
Acquisition Mode : Scan

MS Information

Solvent Delay : 9.00 min

M Absolute : False
M Offset : 0
Resulting EM Voltage : 1905.9

Scan Parameters]

Low Mass : 50.0

Method: MAN_52.M

Fri Aug 04 10:24:41 2006

Page: 2